

**STUDY REPORT W&AR-01
WATER QUALITY ASSESSMENT**

ATTACHMENT B

FIELD QUALITY ASSURANCE DOCUMENTATION

DATE 10-24-08	DOCUMENT#: 19002-00-Series5Sonde
PAGE	REVISION 3

3.0 Series 5, and 5X Sonde Functional Test Data Sheet

Section A:

Service Request # <i>319222835</i>	Customer <i>Rental</i>	Date Started <i>07-25-12</i>
Housing Serial # <i>R49422</i>	Embedded Serial# <i>R49422</i>	Additional Driver Firmware:
Technician <i>HL</i>	Model: Datasonde <i>Minisonde</i> Pipesonde <i>5X</i>	

Customer Display Information

I/D	DOM	Baud Rate	Security	SDI	TTY
Parameter	<i>Time</i>	<i>Temp</i>	<i>pH</i>	<i>Spcond</i>	<i>Sal</i>
Units		<i>°C</i>	<i>units</i>	<i>ms/cm</i>	<i>ppt</i>
Parameter	<i>Turb</i>	<i>100%</i>	<i>100</i>	<i>Depth to</i>	<i>1 part</i>
Units	<i>SC</i>	<i>Sat</i>	<i>mg/l</i>	<i>meters</i>	<i>Volts</i>

For Sonde with Depth - Coefficients

A:	B:	C:	D:
E:	F:	G:	H:
I:	J:	SER:	

FLUOROMETER OFFSETS

1 ST	X10:	X1:
2 ND	X10:	X1:

For Sonde with TDG or PAR - Coefficients

A:	B:	C:	D:
Local:	Ref:		

Performance, Test and Evaluation

Current MPL Rev-- <i>543</i>	pH Electrolyte & Teflon Junction Replaced- Yes	DO membrane Replaced Yes
Upgrade to MPL Rev-- <i>N/A</i>	No NA	No NA
Lenses cleaned - Yes	RTC Battery Replaced Yes	Desiccant Replaced Yes

Section B:

Customer Observations Verified	Submission Day	/ Y / N /	Submission Day	/ Y / N /
	PT&E	PT&E / Upgrade	PT&E	PT&E / Upgrade
Set Time and Date		✓		
Verified all hardware updates as current		✓		
Total current draw. (Circle all that apply)				
MPL PCB 40mA SC Turbidity 20mA				
LDO 70mA 4Beam Turbidity 10mA				
Flurometers:				
1st 30mA 2nd 30mA 3rd 30mA				
PAR 10mA (Optimal Values not to exceed +20mA overall.)				
Current draw of circulator. (20 mA max. beyond previous values.)		120 mA		
Operation of self cleaning motor verified--	P F MA		P F NA	P F NA
Audio functions correctly	(P) F		P F	P F
RTC sleep/wake-up test.	(P) F		P F	P F
Temp probe test at room temperature <i>24.35</i> °C	24.34			
DO 100% sat integrity window verified at +50 mHg over current bp. (Clark Cell only)	P F NA		P F NA	P F NA

DO 100% saturation calibration verified- local - (+/- 0.2 mg/l Clark Cell) (+/- 0.1 mg/l LDO) Scale Factor. (1.5 to 0.5) (LDO only)	<i>X/A</i>		
Conductivity zero (air) calibration verified - (+/- .005mS)	0.0		
Conductivity calibration verified - (+/- .2 mS) 12.856 mS/cm / 47.6 mS/cm	12.86		
Conductivity 1.412mS linearity verified -(+/- .15 mS)	1.396		
Conductivity .100mS verified - (+/- .005 mS)	102		
pH 7 buffer calibration verified- (+/- .2 pH)	N		
pH slope calibration verified at _____ units.			
ORP calibration verified at _____ °C (+/- 20 mV)			
Turbidity - Calibration accepted & verified with DI Water (0.0 +/- 0.7 NTU)			
Turbidity - Calibration accepted & verified at (100.0 +/- 1 NTU) with Hach StabCal	<i>X</i>		
Turbidity - Linearity verified with 40 NTU Hach StabCal - (+/- 4 NTU)			
Depth zero calibration verified - (.02 meters)	0.0		
Depth Check verified - (+/- 0.03 meters)- Tank depth	<i>pws</i>		
Specific Ion Low C mV	High C mV	Specific Ion Low C mV	Specific Ion High C mV
Specific Ion Low C mV	High C mV		High C mV
N03- calibration verified	P F NA	P F NA	P F
NH4+ calibration verified	P F NA	P F NA	P F
Cl- calibration verified	P F NA	P F NA	P F
Chlorophyll 'a' calibration verified	P F NA	P F NA	P F
Rhodamine 'wt' calibration verified	P F NA	P F NA	P F
Blue-green Algae calibration verified	P F NA	P F NA	P F
PAR calibration verified	P F NA	P F NA	P F
TDG calibration verified (+/- 2 mmHg)	P F NA	P F NA	P F
Logging/Sensor Stability Test	P F	P F	P F
pH linearity verified at _____ units: (+/- 0.20 units)			
Battery pack setup and checked	P F NA	P F NA	P F
Hydras3 LT Communications verified, unused slots deactivated.			
Display, Baud Rate, Communications mode settings returned as received.	Yes No	Yes No	Yes No

Calibrated Test Equipment Used -

Description	X-number
Power Supply	X- 7208
DVM Digital Multimeter	X- 72090

Section C. Final Check-off Prior to Submitting for Estimate --

Exterior is clean ✓
Storage cup filled with pH 4, buffer ✓

Hach Business System updated ✓
Date Completed 07-26-12

3.0 Quanta and Quanta G Functional Test Sheet

Section A:

Service Request # <u>31222835</u>	Customer <u>Rental</u>	Serial # <u>QT5747</u>
Technician <u>HL</u>	Model Type <u>Quanta</u> <u>Quanta G</u>	Date Started <u>07-25-12</u>

Performance, Test and Evaluation

Current MPL Rev. <u>3.1</u>	DO membrane installed <u>Yes</u> Yes No NA	PH Electrolyte and Teflon Junction installed - <u>Yes</u> Yes No NA
Lenses cleaned <u>Yes</u> Yes No NA	Desiccant installed <u>Yes</u> Yes No NA	Turbidity Firmware Rev.

Section B:

	Submission Day	Submission Day	Submission Day
Verified customer's observations	Y / N PT&E / Upgrade	Y / N PT&E / Upgrade	Y / N PT&E / Upgrade
Verified proper operation of circulator	(Y) / N / NA	Y / N / NA	Y / N / NA
Temp probe test at room temperature	<u>21.35</u> °C	<u>21.34</u>	
pH 7 Buffer calibration verified (+/- .2 pH)		<u>7.0</u>	
pH slope calibration verified at _____ units.		<u>10.02</u>	
Conductivity calibration verified (+/- .2mS/cm) 12.856 mS/cm / 47.6 mS/cm		<u>12.86</u>	
ORP calibration verified at _____ degrees C (+/- 20 mV)		<u>N/A</u>	
Conductivity .100 mS linearity verified- (+/- .005 mS)		<u>.102</u>	
Conductivity 1.412 mS linearity verified - (+/- .15 mS)		<u>1.413</u>	
DO 100% sat integrity window verified at +50 mmHg above current bp.	(P) / F / NA	P / F / NA	P / F / NA
DO 100% saturation calibration verified - local (+/- .5%)		<u>034</u>	
Turbidity calibration verified in DI water 0.0-(+/- .5NTU)		<u>0.0</u>	
4 Beam Turbidity calibrated at <u>102</u> NTU (+/- 1 NTU) Dilute Formazin verified against "in-house" 2100P		<u>99.8</u>	
4 Beam Turbidity linearity verified at <u>40</u> NTU with Dilute Formazin verified against "in-house" 2100P (+/- 10% of reading.)		<u>40.2</u>	
Depth calibration verified - (+/- .02 meters)		<u>W/T</u>	
Depth check - (+/- .1 meters)			
Logging/Sensor Stability Test	P / F / NA	P / F / NA	P / F / NA
pH linearity verified at _____ units. (+/- 0.20 units)			

Calibrated Test Equipment Used -

Description	X-number
DVM Multimeter	X- <u>AA</u>

Section C. Final Check-off Prior to Submitting for Estimate -

Exterior is clean	<u>✓</u>	Hach Business System updated <u>02 P</u>
Storage cup filled with pH 4 buffer	<u>✓</u>	Date Completed <u>07-26-12</u>

Don Pedro

2012 Sampling and Analysis

DATA REVIEW AND VERIFICATION CHECKLIST

This checklist should be used to document data review verification of data generated through implementation of the FERC-approved study plan.

GENERAL

- For each sample event, samples have been collected and analyzed at all locations and for all analyses specified in the study plan. *Two dry sites*
- For each sample and analyses, the project file contains records field notes, chain-of-custody, and analytical results, including quality assurance documentation (hardcopy and electronic)

FIELD DATA

- Field notes and/or data sheets include date, time of sample collection, field sampling staff, time arrived at site, time left site, site identification, description of site conditions (weather), field parameters, ~~reservoir level or flow information (measured or estimated)~~, sample collection procedures, and call-out quality assurance samples collected. If mistakes are found on the field data sheet, changes can be made by crossing out the mistake and marking the change with a date of change, initials, and reason for change. *N/A not in study plan*
- Documentation of field equipment calibration is in the fieldnotes and/or project records.
- Field data entered into Excel, have been checked by a second-party. *S. Burger*

LABORATORY REPORT

- Field duplicates, blanks, and rinsates were submitted to the laboratory at the frequency specified in the study plan.
- Any constituents found in blanks or rinsates are discussed in the final report.
- Any duplicate concentrations that differ by more than 10% are discussed in the final report.
- Samples were received by the laboratory intact and analyzed within method and/or study specified holding times.
- On laboratory reports, sample IDs, analyses, reporting/detection limits, units, column labels, footnotes, and titles are accurate. Have lab re-issue report with corrections if there are inconsistencies.
- Check that non-detects are always reported in the same manner using consistent notation. For example, either "ND" or "<." Have lab re-issue report with corrections if there are inconsistencies.
- If observed, "J" qualified data and/or elevated detection limits are discussed in the final report.

Table B-1. Rinstate and Trip Blank Water Quality Data--Summer 2012

Analyte	Sample ID	Method	Reporting Limit	FIELD BLANK-1	METHOD BLANK-1	FIELD BLANK-3	METHOD BLANK-3	RINSATE-1	FIELD BLANK-2	METHOD BLANK-2	177261-8	177261-8	
	Date	Detection Limit		8/21/2012	--	8/22/2012	--	8/22/2012	8/23/2012	--	8/23/2012	8/23/2012	
	Sample Type	Field Blank		Method Blank	Field Blank	Method Blank	Rinsate	Field Blank	Method Blank	Original	Duplicate		
	latitude/longitude	Result		Notes	Result	Notes	Result	Notes	Result	Notes	Result	Notes	
	Units	--		737842	4195595	--	--	732763	4183297	732763	4183297	727608	4176308
Basic Water Quality, Inorganic Ions, and Nutrients													
Alkalinity, Total (as CaCO ₃)	mg/L	0.85	1.0	1	ND	1	ND	1	ND	1	ND	12.5	
Ammonia (as N)	mg/L	0.094	0.10	0.1	ND	0.1	ND	0.1	ND	0.1	ND	0.10	ND
Calcium	mg/L	0.0118	0.10	0.0286	J	0.1	ND	0.038	J	0.1	ND	2.83	
Carbon, Dissolved Organic	mg/L	0.021	0.50	0.78	B	0.18	J	0.45	B,J	0.19	J	0.47	B,J
Carbon, Total Organic	mg/L	0.026	0.50	0.34	B,J	0.12	J	0.42	B,J	0.17	J	0.39	B,J
Chloride	mg/L	0.24	1.0	1	ND	1	ND	1	ND	1	ND	1	ND
Hardness, Total	mg/L	0.99	2.0	2	ND	2	ND	2	ND	2	ND	2	ND
Magnesium	mg/L	0.00336	0.10	0.00543	J	0.1	ND	0.00483	J	0.1	ND	0.00361	J
Nitrate (as N)	mg/L	0.037	0.10	0.1	ND	0.1	ND	0.1	ND	0.1	ND	0.047	J
Nitrite (as N)	mg/L	0.016	0.10	0.1	ND	0.1	ND	0.1	ND	0.1	ND	0.10	ND
o-Phosphate (as P)	mg/L	0.031	0.10	0.1	ND	0.1	ND	0.1	ND	0.1	ND	0.10	ND
Phosphorus, Total	mg/L	0.022	0.10	0.037	J	0.1	ND	0.027	J	0.1	ND	0.052	J
Potassium	mg/L	0.103	0.50	0.5	ND	0.5	ND	0.5	ND	0.5	ND	0.5	ND
Sodium	mg/L	0.103	0.50	0.5	ND	0.5	ND	0.5	ND	0.5	ND	0.398	J
Solids, Total Dissolved	mg/L	0.82	1.0	1	ND	1	ND	1	ND	1	ND	1	ND
Solids, Total Suspended	mg/L	0.95	1.0	1	ND	1	ND	1	ND	1	ND	1.0	ND
Total Kjeldahl Nitrogen	mg/L	0.46	0.50	0.5	ND	0.5	ND	0.5	ND	0.5	ND	0.50	ND
Pesticides													
Aldrin	µg/L	0.0016	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Alpha-BHC	µg/L	0.0017	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Beta-BHC	µg/L	0.0039	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Chlordane	µg/L	0.0052	0.025	0.025	ND	0.025	ND	0.025	ND	0.025	ND	0.025	ND
Chlorpyrifos	µg/L	0.0024	0.005	0.0050	ND	0.0050	ND	0.0050	ND	0.0050	ND	0.0050	ND
Delta-BHC	µg/L	0.0016	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Diazinon	µg/L	0.0029	0.0050	0.0050	ND	0.0050	ND	0.0050	ND	0.0050	ND	0.0050	ND
Dieldrin	µg/L	0.0016	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Endosulfan I	µg/L	0.0015	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Endosulfan II	µg/L	0.0016	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Endrin	µg/L	0.0016	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Gamma-BHC	µg/L	0.0023	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Heptachlor	µg/L	0.0018	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Heptachlor Epoxide	µg/L	0.0017	0.010	0.010	ND	0.010	ND	0.010	ND	0.010	ND	0.010	ND
Toxaphene	µg/L	0.023	0.12	0.12	ND	0.12	ND	0.12	ND	0.12	ND	0.12	ND
Total Metals Concentrations													
Arsenic	µg/L	0.04	0.15	0.15	ND	--	J	0.05	--	0.06	J	0.05	J
Cadmium	µg/L	0.003	0.02	0.02	ND	--		0.02	ND	0.003	J	0.02	J
Copper	µg/L	0.01	0.1	0.02	J	--		0.02	J	0.87	--	0.02	J
Iron	µg/L	0.6	10	0.8	J	--		10	ND	2	J	10	ND
Lead	µg/L	0.003	0.04	0.04	ND	--		0.04	ND	0.010	J	0.04	ND
Mercury	ng/L	0.08	0.5	0.5	ND	--		0.16	J	27.4	--	0.08	ND
Methyl Mercury	ng/L	0.026	0.05	0.05	ND	--		0.05	ND	0.436	--	0.05	ND
Selenium	µg/L	0.31	0.6	0.6	ND	--		0.6	ND	0.6	ND	0.6	ND
Silver	µg/L	0.002	0.02	0.02	ND	--		0.02	ND	0.02	ND	0.002	J
Zinc	µg/L	0.03	0.2	0.13	J	--		0.19	J	1.07	--	0.08	J
Dissolved Metals Concentrations													
Arsenic	µg/L	0.04	0.15	0.4	J	--		0.06	J	--		0.27	
Cadmium	µg/L	0.003	0.02	0.02	ND	--		0.02	ND	0.02	ND	0.02	ND
Copper	µg/L	0.01	0.1	0.04	J	--		0.06	J	--		0.45	
Iron	µg/L	0.6	10	10	ND	--		10	ND	0.6	J	10	ND
Lead	µg/L	0.003	0.04	0.04	ND	--		0.04	ND	0.04	ND	0.04	ND
Methyl Mercury	ng/L	0.026	0.05	0.05	ND	--		0.05	ND	0.26	--	0.5	ND
Silver	µg/L	0.002	0.02	0.02	ND	--		0.02	ND	0.02	ND	0.02	ND
Zinc	µg/L	0.03	0.2	0.11	J	--		0.20	--	0.91	--	0.09	J

-- Laboratory methods do not include method blanks specific to water quality study metals analyses.

B Analyte was present in the associated method blank.

FB Field Blank

J Analyte was detected at a concentration below the reporting limit and above the laboratory method detection limit. Reported value is estimated.

ND Analyte included in the analysis, but not detected at the reporting limit.